

Validation Addendum 1 Statistical Report

Method: VSDVAC 66 Version 0.00, An ELISA Method for the Detection of IgG Specific to SARS-CoV-2 Nucleocapsid Protein in Human Serum

PPD Project: RPPF2

Reagent ID: VSDVAC 66 VS Adden1 (b) (4) 01

Qualification of a New (b) (4) (Lot (b) (4)) for Use in the Nucleocapsid IgG ELISA

Version: 1.0

Prepared for Moderna

by PPD® Laboratories 2244 Dabney Road Richmond, Virginia 23230 (804) 359-1900

Original Document Date: 08 December 2020 Revision Date: NA

Confidentiality Statement

The information in this document contains trade secrets and commercial information that are privileged or confidential and may not be disclosed unless such disclosure is required by applicable law or regulations. In any event, persons to whom the information is disclosed must be informed that the information is privileged or confidential and may not be further disclosed by them. These restrictions on disclosure will apply equally to all future information supplied to you which is indicated as privileged or confidential.

		Confidential
PPD S	Signatures	
Statistical Author:	8	
Victoria Pisciella	Victoria A. Pisciella Senior Biostatistician II I am the author of this docume 14 Dec 2020 09:49:24 -05:00	nt Docu <i>Sign</i> ,
Victoria Pisciella	D	ate
Second-Statistician Reviewer: (b) (6)	(b) (6) (b) (6) I reviewed this document 14 Dec 2020 11:03:55 -05:00	Docu <i>Sig</i> r
(b) (6)	D	ate
Scientific Contributions By:	Carl Breidenbach Associate Group Leader I approve this document 14 Dec 2020 09:41:09 -05:00	Docu <i>Sig</i> n
Carl Breidenbach		ate
Scientific Reviewers and Approvers:	Adrienne Howlett Manager Labs I approve this document 14 Dec 2020 09:58:27 -05:00	Docu <i>Sig</i> n.
Adrienne Howlett Southernell	Marie Bonhomme Associate Director I approve this document 14 Dec 2020 09:52:04 -05:00	ate DocuSign
Marie Bonhomme, Ph.D.	D	ate
Quality Assurance Reviewer: (b) (6)	(b) (6) (b) (6) I approve this document 14 Dec 2020 09:56:23 -05:00	Docu <i>Sign</i> ,
(b) (6)	D	ate
cc List: Tina Green; Christopher Hammond; (b) (6)		


			Confidential
	Mode	erna Approval	
Client approval of this rep PPD's ECM system.	ort may be granted in	the form of written communicat	ion and will be stored in
Scientific Approver:			
(b)	(6)	(b) (6) (b) (6) Toonfirm that this approval was given via email and housed in E 14 Dec 2020 09:29:14 -05:00	CM under the project code(s) listed in the plan.
			Docu <i>Sign</i>
Rolando Pajon, Ph.D.			ate
Director, Clinical Bioma	rkers		

Confidential

Experiment Summary


Background: The currently qualified lot of (b) (4) (Lot (b) (4)), which is used in the Nucleocapsid IgG ELISA, is nearly depleted. For this reason, an experiment was performed to qualify a new lot of (b) (4) (Lot (b) (4)) as a replacement for the current lot. The objective of this statistical report is to assess whether the new lot produces assay results that are acceptably similar to those produced when using the qualified lot.

Experimental Design: Per the Validation Plan Addendum 1^[1], the new lot and the qualified lot were tested



(b) (4) All testing was performed following VSDVAC 66 Version 0.00^[3].

<u>Statistical Results:</u> Sample antibody concentrations are provided in <u>Attachment I</u>. QCS antibody concentrations are displayed graphically in <u>Attachment II</u>. The experimental results and their respective pre-specified acceptance criteria are summarized in the table below.



Conclusion

All pre-specified acceptance criteria were met. Therefore, the new working concentration of (b) (4) (b) (4) is qualified for use in the Nucleocapsid IgG ELISA.



Assay Information

The quantitative ELISA (referred to as Nucleocapsid IgG ELISA) was designed to detect IgG antibody to the SARS-CoV-2 virus in human serum. (b) (4)

(b) (4)

(b) (4)

Statistical Results


Statistical methods utilized to analyze data within this experiment are provided in the Statistical Methods Reference Guide^[4].


Plate Status:

One plate failed to meet the plate acceptance criteria detailed in the Validation Statistical Report^[5]. Run Plate Plate tested with the new lot, failed due to too few valid standard curve points. The plates were repeated pairwise, qualified versus new, and the repeat results replaced the original results in all analyses. The inclusion status for each plate is provided in Table 2.

Standard Curves:

The standard curve fits for each of the runs are graphically depicted in Attachment III. As shown in Attachment III, the (b) (4) are in line with those determined during assay validation^[5].





Scientific Contribution

During review of experimental Run and Run (b) (4) Repeat (c) it was found that the preparation of the (b) (4) for use with the candidate(b) (4) followed an incorrect dilution factor. It was noted upon review of experimental Run (c) Repeat (c) the plate was significantly overdeveloped, resulting in elevated background and elevated responses across the plate. Due to these noted technical errors, these three runs were rejected and repeat experiments performed. All pre-specified acceptance criteria were met and the (b) (4) lot (b) (4) is considered acceptable for use in the assay at a working concentration of (b) (4) using a (b) (4)

The validation plan addendum listed the antigen lot number as (b) (4) as per the Certificate of Analysis (CoA) from (b) (4) The vendor, (b) (4) clarified this was an error. The lot number should be listed as (b) (4)

It was noted that there was variation in the standard curves, especially within the qualified lot (b) (4); however, all runs used in statistical analysis passed validity criteria. Standard curves analyzed on plates using the new lot of (b) (4) produced more consistent results across plates when compared to plates coated with the qualified (b) (4) All runs used in statistical analysis passed method acceptance criteria established during the validation. It is recommended that standard curve performance be monitored during routine testing to assess variability.

		Confidential
--	--	--------------

References

- [1] Validation Plan Addendum 1: Addendum to the Validation of an ELISA Method for the Detection of IgG Specific to SARS-CoV-2 Nucleocapsid Protein in Human Serum, RPPF2, VSDVAC_66_VP_Addend1, 21-October-2020.
- [2] Validation Plan Addendum 1 Amendment 1: Amendment to the Addendum to the Validation of an ELISA Method for the Detection of IgG Specific to SARS-CoV-2 Nucleocapsid Protein in Human Serum, RPPF2, VSDVAC 66 VP Addend1 Amend1, 02-November-2020.
- [3] VSDVAC 66: An ELISA Method for the Detection of IgG Specific to SARS-CoV-2 Nucleocapsid Protein in Human Serum, v0.00.
- [4] PPD Nonclinical Statistics: Statistical Methods Reference Guide v1.0, 21-December-2009.
- [5] Validation Statistical Report: Validation of an ELISA Method for the Detection of IgG Specific to SARS-CoV-2 Nucleocapsid Protein in Human Serum, RPPF2, v1.0, 30-November-2020.

[6]	# N / A	(b) (4)	
	(b) (4)		

|--|

Figures, Tables, and Attachments

Table 1a Reagent Details




Table 1b Sample Information




Table 2 Experimental Design




Figure 1




Table 3







Table 5





Attachment II QCS Antibody Concentrations (AU/mL)



Attachment III

